



MISSOURI DEPARTMENT OF HEALTH
STATE PUBLIC HEALTH LABORATORY
DATAMASTER MAINTENANCE REPORT

BREATH ALCOHOL PROGRAM

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN

940092

DATE OF INSPECTION

JUN -8 2009

LOCATION OF INSTRUMENT (STREET AND CITY)

100 N. MULBERRY FAYETTE MO 65248

TIME OF INSPECTION

1339

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED)

☒ COMPUTER

☒ DETECTOR

☒ PROGRAM

☒ FILTERS

☒ HEATERS SAMPLE CHAMBER 49 °C

☒ QUARTZ STANDARD

☒ FLOW DETECTOR

☒ CALIBRATION

☒ PUMP HIGH SPEED

☒ PRINTER

☒ INDICATOR LIGHTS

☒ TIME AND DATE 06/02/09 1339

☒ SIMULATOR TEMPERATURE (34 °C ± 0.2 °C) 34 °C

☒ CALIBRATION CHECK -

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 .098

TEST 2 .099

TEST 3 .099

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

☒ NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (Over .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

INSTRUMENT OPERATING WITHIN DEPT.

OF HEALTH REGULATIONS.

REPCO MARKETING SOLUTION LOT # 08001

EXP. DATE 2/11/09 SOLUTION .10 %

INSPECTING OFFICER

SIGNATURE

GREG T LANHAM

PRINT NAME

GREG T LANHAM

TYPE II PERMIT NUMBER/EXPIRATION DATE

820317 / 11/21/10

TELEPHONE NUMBER

660-248-2241

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 08001

EXPIRATION DATE: August 11, 2009 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 08001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed utilizing a gas chromatograph and found to contain 0.1216 gms/dl wt/vol ethyl alcohol in aqueous solution (i.e. ethanol).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of 0.100% +/-3% when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator.

The date of manufacture for this lot number is February 12, 2008.
The expiration date for this lot number is August 11, 2009 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
FAYETTE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940092
06/02/09

TESTING OFFICER:

LANHAM/GREG-T

OFFICER T.I.: 3

PERMIT NUMBER: 020017

EXPIRATION DATE: 11/21/19

MISCELLANEOUS DATA:

--- SUPERVISOR CODE ---

BLANK TEST	.000	13:59
INTERNAL STANDARD	VERIFIED	10:59
EXTERNAL STANDARD	.000	14:00
BLANK TEST	.000	14:00
EXTERNAL STANDARD	.000	14:01
BLANK TEST	.000	14:02
EXTERNAL STANDARD	.000	14:02
BLANK TEST	.000	14:02

N = 3

SM = .1

AND = .0000

Operator Signature

Greg T. Lanham #3

Printed on recycled paper with agri-based inks

CMSU 22

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
FAYETTE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940092
06/02/09
13:59

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

HEATERS
SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~

Operator Signature

Greg T. Lanham #3

Printed on recycled paper with agri-based inks

CMSU 2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
FAYETTE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940910
06/02/09

ARREST TIME: 10:00

SUBJECT NAME:

SAMPLE

DOB: 10/03/90 SEX: F

STATE/D.L.: MO-000000000

ARRESTING OFFICER:

LANHAM/DREB/T

OFFICER I.D.: 2

TESTING OFFICER:

LANHAM/DREB/T

OFFICER I.D.: 3

PERMIT NUMBER: 820017

EXPIRATION DATE: 11/01/10

MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST 1.000 10:40

INTERNAL STANDARD VERIFIED 10:45

RADIO INTERFERENCE

Operator Signature

Brec T Lanham 10/10

Printed on recycled paper with agri-based inks

CMSU 2208-02

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



GREG T LANHAM

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s)

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 11/21/08
Number 820317
Expires 11/21/2010

MO 580-0771 (7-88)

John J Mathewson

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)